

Fully Autoclavable Fixed Volume Micropipette AFQ-2000



Fully Autoclavable Fixed Volume Micropipette AFQ-2000 is fully autoclavable, without disassemble or dismantle. Used to dispense sample measurement between 2000 µl volume ranges.

Additional features maintenance-free seals, integral tip ejector for access into narrow tubes: single-channel pipettors, calibrated in accordance ISO/DIS 8655 and each pipettor passes a quality control calibration test and is supplied with a certificate and feature Fully autoclavable, portable ergonomic design, clear volume display, Easy and simple calibration, avoid repetitive strain, provide individual test certificate ISO 8655/DIN 12650 with a limited 3-year warranty.

Technical Specification:

Model	AFQ-2000
Volume	2000 µl
Precision	0.3%
Accuracy ±%	0.2%
Catalogue No.	4001555-04

Technical Data:

Model	Volume (µl)	Test Volume	Inaccuracy	Imprecision	Catalogue No.
AFQ-5	5µl	5µl	1.3%	1.2%	4001505-04
AFQ-10	10µl	10µl	0.8%	0.8%	4001510-04
AFQ-20	20µl	20µl	0.6%	0.5%	4001515-04
AFQ-25	250µl	250µl	0.5%	0.3%	4001520-04
AFQ-50	50µl	50µl	0.5%	0.3%	4001525-04
AFQ-100	100µl	100µl	0.5%	0.3%	4001530-04
AFQ-200	200µl	200µl	0.4% 0	.2%	4001535-04
AFQ-250	250µl	250µl	0.4%	0.2%	4001540-04
AFQ-500	500µl	500µl	0.3%	0.2%	4001545-04
AFQ-1000	1000µl	1000µl	0.3%	0.2%	4001550-04
AFQ-2000	2000µl	2000µl	0.3%	0.2%	4001555-04
AFQ-5000	5000µl	5000µl	0.3%	0.15%	4001560-04



Iris Analytical Ltd.

66 W Flagler St #900 Miami, Florida 33130 United States